

# TEST REPORT



ORIGINAL

**Test Report No.** 70.405.20.13797.01R1  
**Dated** 2020-05-14


**Applicant:** Wuxi ATA Pack Co., Ltd  
**Address:** 286-901/No 286-901, North Changjiang Road, Xinwu District, Wuxi City, Jinagsu Province  
**Attn:** Yang Liqiang

**Order No.:** -  
**Style No.:** Universal  
**Season:** All Seasons  
**No. of samples:** -  
**Buyer:** -  
**Country of Origin:** China  
**Goods Export To:** USA

**Care Label Provided:** -

**Receipt Date of Sample:** Received on 2020-05-13  
**Date of Testing:** From 2020-05-13 To 2020-05-13  
**Sample Submitted:** The sample(s) was (were) submitted by applicant and identified.  
The product's function is declared by the applicant.  
The testing point(s) was (were) selected by the applicant.

**Test Result:** Refer to Next Page.



TÜV SÜD Certification and Testing (China) Co., Ltd. Shanghai Branch  
Testing Center  
Prepared by:

*Andy Wang*

Andy Wang  
Softlines Department

Authorized by:

*Sherry Wang*

Sherry Wang  
Softlines Department

Note: (1) The TÜV SÜD Certification and Testing (China) Co., Ltd. "General Terms & Conditions" applied.  
For full version, please visit: <http://www.tuv-sud.cn/cn-scn/terms-and-conditions>  
(2) The results relate only to the Items tested, (3) The test report shall not be reproduced except in full without the written approval of the laboratory.  
(4) Any use for advertising purposes must be granted in writing. This technical report may only be quoted in full. This report is the result of a single examination of the object in question and is not generally applicable evaluation of the quality of other products in regular production. For further details, please see testing and certification regulation, chapter A-3.4.

Laboratory:  
TÜV SÜD Certification and Testing  
(China) Co., Ltd. Shanghai Branch  
Testing Center Softlines Lab

Phone: +86 21 6037 6300  
Fax: +86 21 6037 6350  
E-mail: [softlines@tuv-sud.cn](mailto:softlines@tuv-sud.cn)  
[www.tuv-sud.cn/softlines](http://www.tuv-sud.cn/softlines)

Regd. Office:  
TÜV SÜD Certification and Testing (China) Co., Ltd. Shanghai  
Branch Testing Center Softlines Lab  
2F Building B, No. 1999 Du Hui Rd, Minhang District, Shanghai  
201108 P.R. China

Page 1 of 2


# TEST REPORT

Test Report No. 70.405.20.13797.01R1  
Dated 2020-05-14



ORIGINAL

## Description of The Test Subject

Sample	Description	Photo
001	CPE Gown in blue for waterproof and anti-static	

## Test Results

### 1 Water Resistance-Impact Penetration Test

AATCC 42:2017e

Weight of blotter gained(g)

001-Front panel

0.0

### 2 Hydrostatic head Test

AATCC 127:2017(2018)e Option 2; Hydrostatic Head; Rate Of Increase Of Water Pressure: 60cm/min.,  
Face Side Facing Water

As Received  
(cm.H<sub>2</sub>O)

001-Front panel

76.5

Remark: This Report is to Supersede Test Report No.70.405.20.13797.01 Issued on 2020-05-13.

-End of The Test Report-